

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 07/01/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21100

*RMS
 21-25675-*

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:19
INTERNAL STANDARD	VERIFIED	13:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:19
BLANK TEST	0.00	13:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:20
BLANK TEST	0.00	13:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:21
BLANK TEST	0.00	13:22

