

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 03/30/2021

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20500

*RMS
21-11281*

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:11
INTERNAL STANDARD	VERIFIED	12:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:11
BLANK TEST	0.00	12:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:12
BLANK TEST	0.00	12:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:13
BLANK TEST	0.00	12:14

