

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 04/06/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21020

*RMS
 21-17302*

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:24
INTERNAL STANDARD	VERIFIED	13:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:24
BLANK TEST	0.00	13:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:25
BLANK TEST	0.00	13:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:27
BLANK TEST	0.00	13:27

