

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

CM
21-25338

Date: 06/29/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:31
INTERNAL STANDARD	VERIFIED	12:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:32
BLANK TEST	0.00	12:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:33
BLANK TEST	0.00	12:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:34
BLANK TEST	0.00	12:35

