

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 08/31/2021

RM
21-34269

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:59
INTERNAL STANDARD	VERIFIED	10:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:59
BLANK TEST	0.00	11:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:00
BLANK TEST	0.00	11:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:01
BLANK TEST	0.00	11:02

