

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 10/19/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21280

RMS
21-1/14/99

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:30
INTERNAL STANDARD	VERIFIED	11:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:31
BLANK TEST	0.00	11:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:32
BLANK TEST	0.00	11:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:33
BLANK TEST	0.00	11:34

