

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 02/02/2021

*RMS*  
*21-4012*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:05
INTERNAL STANDARD	VERIFIED	15:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:06
BLANK TEST	0.00	15:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:07
BLANK TEST	0.00	15:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:08
BLANK TEST	0.00	15:09

