

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

RW
21-27314

Date: 07/13/2021

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:48
INTERNAL STANDARD	VERIFIED	12:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:48
BLANK TEST	0.00	12:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:49
BLANK TEST	0.00	12:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:50
BLANK TEST	0.00	12:51

