

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 09/07/2021

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 21170

*RMS*  
*21-35297*

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:09
INTERNAL STANDARD	VERIFIED	12:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:10
BLANK TEST	0.00	12:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:11
BLANK TEST	0.00	12:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:12
BLANK TEST	0.00	12:13

