

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
22-41190

Date: 10/12/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:10
INTERNAL STANDARD	VERIFIED	15:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:11
BLANK TEST	0.00	15:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:12
BLANK TEST	0.00	15:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:13
BLANK TEST	0.00	15:14

