

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
22-18404

Date: 05/17/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:34
INTERNAL STANDARD	VERIFIED	11:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:34
BLANK TEST	0.00	11:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:35
BLANK TEST	0.00	11:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:36
BLANK TEST	0.00	11:37

