

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

*RMS*  
*22-22761*

Date: 06/14/2022

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:53
INTERNAL STANDARD	VERIFIED	13:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:53
BLANK TEST	0.00	13:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:55
BLANK TEST	0.00	13:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:56
BLANK TEST	0.00	13:57

