

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 08/23/2022

RMS
22-33796

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:15
INTERNAL STANDARD	VERIFIED	13:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:15
BLANK TEST	0.00	13:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:16
BLANK TEST	0.00	13:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:17
BLANK TEST	0.00	13:18

