

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 10/18/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22260

ZMS
22-12073

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:15
INTERNAL STANDARD	VERIFIED	13:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:15
BLANK TEST	0.00	13:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:17
BLANK TEST	0.00	13:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:18
BLANK TEST	0.00	13:19

