

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 03/01/2022

RMS
22-7439

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:13
INTERNAL STANDARD	VERIFIED	13:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:14
BLANK TEST	0.00	13:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:15
BLANK TEST	0.00	13:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:16
BLANK TEST	0.00	13:17

