

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

*RMS*  
*22-72549*

Date: 08/11/2022

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 84622  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 22109

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:07
INTERNAL STANDARD	VERIFIED	09:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:08
BLANK TEST	0.00	09:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:09
BLANK TEST	0.00	09:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:10
BLANK TEST	0.00	09:11

