

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 09/27/2022

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22190

ZMS
22-38930

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:01
INTERNAL STANDARD	VERIFIED	13:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:01
BLANK TEST	0.00	13:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:02
BLANK TEST	0.00	13:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:03
BLANK TEST	0.00	13:04

