

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
23-8096

Date: 02/21/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:29
INTERNAL STANDARD	VERIFIED	14:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:29
BLANK TEST	0.00	14:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:30
BLANK TEST	0.00	14:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:31
BLANK TEST	0.00	14:32

