

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
23-9295

Date: 03/01/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:12
INTERNAL STANDARD	VERIFIED	14:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:13
BLANK TEST	0.00	14:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:14
BLANK TEST	0.00	14:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:15
BLANK TEST	0.00	14:16

