

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 05/16/2023

RMS
23-22080

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:32
INTERNAL STANDARD	VERIFIED	10:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:32
BLANK TEST	0.00	10:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:33
BLANK TEST	0.00	10:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:34
BLANK TEST	0.00	10:35

