

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 08/01/2023

LMS
23-31765

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:57
INTERNAL STANDARD	VERIFIED	10:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:57
BLANK TEST	0.00	10:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:58
BLANK TEST	0.00	10:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:59
BLANK TEST	0.00	11:00

