

SOLUTION CHANGE

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

*RMJ*  
*23-49563*

Date: 11/02/2023

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:51
INTERNAL STANDARD	VERIFIED	12:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:51
BLANK TEST	0.00	12:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:52
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:54

