

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RMS
23-51235

Date: 11/14/2023

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:54
INTERNAL STANDARD	VERIFIED	11:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:55
BLANK TEST	0.00	11:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:56
BLANK TEST	0.00	11:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:57
BLANK TEST	0.00	11:58

