

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 01/10/2023

*RM,
23-1490*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:49
INTERNAL STANDARD	VERIFIED	11:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:49
BLANK TEST	0.00	11:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:50
BLANK TEST	0.00	11:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:51
BLANK TEST	0.00	11:52

