

SOLUTION CHANGE

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

*RM*  
*23-4904*

Date: 02/01/2023

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:30
INTERNAL STANDARD	VERIFIED	14:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:31
BLANK TEST	0.00	14:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:32
BLANK TEST	0.00	14:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:33
BLANK TEST	0.00	14:34

