

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

*RMS*  
*23-14255*

Date: 04/05/2023

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:42
INTERNAL STANDARD	VERIFIED	09:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:42
BLANK TEST	0.00	09:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:43
BLANK TEST	0.00	09:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:44
BLANK TEST	0.00	09:45

