

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 05/30/2023

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23030

*RMS*  
*23-21230*

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 11:58 |
| INTERNAL STANDARD       | VERIFIED | 11:58 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 11:58 |
| BLANK TEST              | 0.00     | 11:59 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 11:59 |
| BLANK TEST              | 0.00     | 12:00 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 12:00 |
| BLANK TEST              | 0.00     | 12:01 |

