

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 06/20/2023

*12MS
23-27741*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:21
INTERNAL STANDARD	VERIFIED	12:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:21
BLANK TEST	0.00	12:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:22
BLANK TEST	0.00	12:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:23
BLANK TEST	0.00	12:24

