

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

RMS
23-32415

Date: 07/18/2023

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:51
INTERNAL STANDARD	VERIFIED	10:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:51
BLANK TEST	0.00	10:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:52
BLANK TEST	0.00	10:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:53
BLANK TEST	0.00	10:54

