

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 08/29/2023

*KMS*  
*23-39491*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:32
INTERNAL STANDARD	VERIFIED	12:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:32
BLANK TEST	0.00	12:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:33
BLANK TEST	0.00	12:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:34
BLANK TEST	0.00	12:35

