

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RMS
23-44115

Date: 09/26/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:13
INTERNAL STANDARD	VERIFIED	12:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:14
BLANK TEST	0.00	12:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:15
BLANK TEST	0.00	12:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	12:16
BLANK TEST	0.00	12:17

