

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 11/28/2023

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23270

*LMS  
23-53709*

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:22
INTERNAL STANDARD	VERIFIED	11:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:22
BLANK TEST	0.00	11:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	11:24
BLANK TEST	0.00	11:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	11:25
BLANK TEST	0.00	11:26

