

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 02/27/2024

LMS
24-8579

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:18
INTERNAL STANDARD	VERIFIED	12:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	12:18
BLANK TEST	0.00	12:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	12:19
BLANK TEST	0.00	12:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	12:20
BLANK TEST	0.00	12:21

