

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 04/23/2024

LMJ
29-17715

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:22
INTERNAL STANDARD	VERIFIED	11:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:22
BLANK TEST	0.00	11:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:23
BLANK TEST	0.00	11:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:24
BLANK TEST	0.00	11:25

