

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 01/20/2021

*RMS  
21-235-3*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:57
INTERNAL STANDARD	VERIFIED	12:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:57
BLANK TEST	0.00	12:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:58
BLANK TEST	0.00	12:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:59
BLANK TEST	0.00	13:00

