

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 09/21/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21170

RMS
21-37320

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:52
INTERNAL STANDARD	VERIFIED	12:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:54
BLANK TEST	0.00	12:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:55
BLANK TEST	0.00	12:56

