

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 05/25/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21020

RMS
21-19871

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:35
INTERNAL STANDARD	VERIFIED	12:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:35
BLANK TEST	0.00	12:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:37
BLANK TEST	0.00	12:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:38
BLANK TEST	0.00	12:38

