

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 10/26/2021

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 21280

*RMS*  
*21-4/24/97*

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:52
INTERNAL STANDARD	VERIFIED	12:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:53
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:54
BLANK TEST	0.00	12:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:55
BLANK TEST	0.00	12:56

