

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 03/01/2022

Rm
22-7139

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:16
INTERNAL STANDARD	VERIFIED	13:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:17
BLANK TEST	0.00	13:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:18
BLANK TEST	0.00	13:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:19
BLANK TEST	0.00	13:20

