

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 01/18/2022

*12MS
 22-2184*

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:43
INTERNAL STANDARD	VERIFIED	13:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:43
BLANK TEST	0.00	13:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:44
BLANK TEST	0.00	13:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:45
BLANK TEST	0.00	13:46

