

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 05/10/2022

*RM*  
 22-17209

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:10
INTERNAL STANDARD	VERIFIED	15:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:11
BLANK TEST	0.00	15:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:12
BLANK TEST	0.00	15:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:13
BLANK TEST	0.00	15:14

