

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 10/25/2022

RMS
22-1/3084

OPERATOR NAME:
 MICHEAL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:22
INTERNAL STANDARD	VERIFIED	09:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:22
BLANK TEST	0.00	09:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:24
BLANK TEST	0.00	09:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:25
BLANK TEST	0.00	09:25

