

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 12/20/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22390

RMS
22-51315

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:03
INTERNAL STANDARD	VERIFIED	13:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:03
BLANK TEST	0.00	13:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:04
BLANK TEST	0.00	13:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:05
BLANK TEST	0.00	13:06

