

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 02/01/2022

RMS
22-3983

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:29
INTERNAL STANDARD	VERIFIED	13:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:29
BLANK TEST	0.00	13:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:30
BLANK TEST	0.00	13:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:31
BLANK TEST	0.00	13:32

