

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 06/03/2022

RM >
22-2110

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:03
INTERNAL STANDARD	VERIFIED	11:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:03
BLANK TEST	0.00	11:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:05
BLANK TEST	0.00	11:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:06
BLANK TEST	0.00	11:06

