

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RMS
22-15198

Date: 11/08/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:57
INTERNAL STANDARD	VERIFIED	09:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:58
BLANK TEST	0.00	09:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:00
BLANK TEST	0.00	10:01

