

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

*ZMS*  
*22-8373*

Date: 03/08/2022

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:02
INTERNAL STANDARD	VERIFIED	13:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:02
BLANK TEST	0.00	13:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:03
BLANK TEST	0.00	13:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:04
BLANK TEST	0.00	13:05

