

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 06/28/2022

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:28
INTERNAL STANDARD	VERIFIED	14:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:29
BLANK TEST	0.00	14:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:30
BLANK TEST	0.00	14:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:31
BLANK TEST	0.00	14:32

RM
22-24988

