

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RMS
22-26077

Date: 07/05/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:08
INTERNAL STANDARD	VERIFIED	13:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:08
BLANK TEST	0.00	13:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:09
BLANK TEST	0.00	13:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:10
BLANK TEST	0.00	13:11

