

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 09/20/2022

ZMS
22-37971

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:26
INTERNAL STANDARD	VERIFIED	14:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:26
BLANK TEST	0.00	14:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:27
BLANK TEST	0.00	14:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:29
BLANK TEST	0.00	14:29

